

FEB - 9 2001

510(k) Summary of Safety & Effectiveness

K010179

This 510(k) Summary of Safety and Effectiveness for the EBI VueCath™ Spinal Endoscopic System is provided as required per Section 513(I)(3) of the Food, Drug and Cosmetic Act.

1. **Submitter:** EBI, L.P.
100 Interpace Parkway
Parsippany, NJ 07054

Contact Person: Patricia Flood
Tel: (973) 299-9022, x3318

Date prepared: January 18, 2001

2. **Proprietary Name:** EBI VueCath™ Spinal Endoscopic System
Common Name: Catheter
Classification Name: Catheter, Conduction, Anesthesia (868.5120)

3. **Predicate or legally marketed* devices that are substantially equivalent:**

- EBI VueCath™ Spinal Endoscopic System (EBI, L.P.)
- Myelotec Myeloscope System – Myelotec, Inc.

4. **Description of the device:** The EBI VueCath™ Spinal Endoscopic System is an arthroscope consisting of several components and different accessories for viewing the lumbar and sacral spinal anatomy. This system includes a fiberscope, disposable catheter, and various accessories. The fiberscope is designed to connect to any compatible commercially available endoscopic video imaging system by using a camera coupler and light cord adapters.

5. **Intended Use:** When used with a fiberoptic endoscope, the EBI VueCath™ Spinal Endoscopic System can be used in the lumbar and sacral spine for observing epidural anatomy, pathology, and delivery of drugs approved for epidural indications.

In addition, the system may be used by physicians for the illumination and visualization of tissues of the epidural space in the lumbar and sacral spine for the purpose of assisting in the diagnosis of disease utilizing a caudal approach via the sacral hiatus.

6. **Materials:** The catheter is the patient contacting portion of the system. It is manufactured from medical grade polyurethane. The catheter shaft contains barium sulfate for radiopacity and titanium dioxide as a whitening colorant. This 510(k) is being submitted for the addition of these two materials to the catheter tip. The outer jacket of the fiberscope is made out of polyimide. The HaloScope contains a carbon black colorant.

- 7. Comparison of the technological characteristics of the device to predicate devices:** There are no significant differences between the modified EBI VueCath™ Spinal Endoscopic System and other spinal arthroscopes currently on the market. It is substantially equivalent* to the predicate devices in design, materials and intended use. Also, mechanical testing demonstrates that the device meets its functional requirements.

*Any statement made in conjunction with this submission regarding a determination of substantial equivalence to any other product is intended only to relate to whether the product can be lawfully marketed without pre-market approval or reclassification and is not intended to be interpreted as an admission or any other type of evidence in patent infringement litigation. [Establishment Registration and Premarket Notification Procedures, Final Regulation, Preamble, August 23, 1977, FR 42520 (Docket No. 76N-0355.)]



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

FEB - 9 2001

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Mr. Jon Caparotta, RAC
Manager, Regulatory Affairs
EBI, L.P.
100 Interpace Parkway
Parsippany, New Jersey 07054

Re: K010179
Trade Name: EBI VueCath™ Spinal Endoscopic System
Regulatory Class: II
Product Code: HRX
Dated: January 18, 2001
Received: January 19, 2001

Dear Mr. Caparotta:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895.

A substantially equivalent determination assumes compliance with the current Good Manufacturing Practice requirement, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic (QS) inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4595. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its internet address "<http://www.fda.gov/cdrh/dsmamain.html>".

Sincerely yours,

for Miriam C. Provost
Celia M. Witten, Ph.D., M.D.
Director
Division of General, Restorative
and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

STATEMENT OF INDICATIONS FOR USE

Page 1 of 1

510(k) Number (if known): K010179

Device Name: EBI VueCath™ Spinal Endoscopic System

Indications For Use:

When used with a fiberoptic endoscope, the EBI VueCath™ Spinal Endoscopic System can be used in the lumbar and sacral spine for observing epidural anatomy, pathology, and delivery of drugs approved for epidural indications.

In addition, the system may be used by physicians for the illumination and visualization of tissues of the epidural space in the lumbar and sacral spine for the purpose of assisting in the diagnosis of disease utilizing a caudal approach via the sacral hiatus.

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use ☒
(Per 21 CFR 801.109)

OR

Over-The-Counter Use
(Optional Format 1-2-96)

Miriam C. Provost
(Division Sign-Off)
Division of General, Restorative
and Neurological Devices

510(k) Number K010179